



February 23, 2021

To Whom it May Concern:

The KAMRA Inlay is considered **MR Conditional**. The device is non-metallic and non-magnetic.

MR Conditional items demonstrate safety in the MR environment within defined conditions including conditions for the static magnetic field, the time-varying gradient magnetic fields, and the radio frequency fields.

MRI Safety Information

The KAMRA Inlay is MR Conditional. A patient with this device can be safely scanned in an MR system meeting the following conditions:

Static magnetic field	3.0 T or less
Maximum magnetic field spatial gradient	No limit
RF Excitation	Circularly Polarized (CP)
RF Transmit Coil Type	Whole body transmit coil, Head RF transmit-receive coil or Extremity RF transmit-receive coil
Maximum Whole Body SAR	4 W/kg
Limits on Scan Duration	1 hour
MR Image Artifact	The presence of the implant may produce an image artifact

Sincerely,

DocuSigned by:

A handwritten signature in black ink that reads "Scott Carlisle". The signature is enclosed in a blue DocuSign signature box.

DBA11E3200B8471...

Scott Carlisle
Director QA and Device Manufacturing